

Implementation of a Coded Aperture Imaging System for Gamma Measurement and Experimental Feasibility Tests

Kwangdon Kim¹, Hakjae Lee², Jinwook Jang³, Yonghyun Chung⁴, Donghoon Lee⁴, Chanwoo Park⁴, Jinhun Joung^{3,5}, Yongkwon Kim^{3,5}, and Kisung Lee³

¹ Department of IT Convergence, Korea University, Seoul, Korea photon@korea.ac.kr

² Research Institute of Global Health Tech., Korea University / Seoul, Korea tomato98@korea.ac.kr

³ Department of Bio-convergence Engineering, Korea University, Seoul, Korea {kisung, cs3132}@korea.ac.kr

⁴ Department of Radiological Science, Yonsei University, Wonju, Korea {ychung, ldhdaum, cksdn9432}@yonsei.ac.kr

⁵ Nucare Medical System Inc., Seoul, Korea {jinhun.joung, yongkwon.kim}@nucaremed.com

* Corresponding Author: Kisung Lee, kisung@korea.ac.kr

Received January 18, 2017; Accepted January 27, 2017; Published February 28, 2017

* Short Paper

Abstract: Radioactive materials are used in medicine, non-destructive testing, and nuclear plants. Source localization is especially important during nuclear decommissioning and decontamination because the actual location of the radioactive source within nuclear waste is often unknown. The coded-aperture imaging technique started with space exploration and moved into X-ray and gamma ray imaging, which have imaging process characteristics similar to each other. In this study, we simulated 21x21 and 37x37 coded aperture collimators based on a modified uniformly redundant array (MURA) pattern to make a gamma imaging system that can localize a gamma-ray source. We designed a 21x21 coded aperture collimator that matches our gamma imaging detector and did feasibility experiments with the coded aperture imaging system. We evaluated the performance of each collimator, from 2 mm to 10 mm thicknesses (at 2 mm intervals) using root mean square error (RMSE) and sensitivity in a simulation. In experimental results, the full width half maximum (FWHM) of the point source was 5.09° at the center and 4.82° at the location of the source was 9°. We will continue to improve the decoding algorithm and optimize the collimator for high-energy gamma rays emitted from a nuclear power plant.

Keywords: Coded aperture, Gamma ray imaging, Image & video sensing and acquisition

1. Introduction

Radioactive materials are used in medicine, non-destructive testing, and nuclear plants. If the materials are not accurately managed or are badly located, radiation can cause personal damage and health problems. Source localization is especially important during nuclear decommissioning and decontamination, because the actual location of the radioactive source within nuclear waste is often unknown [1, 2].

The coded-aperture imaging technique started with space exploration and then moved into X-ray and gamma ray imaging, which have imaging process characteristics similar to each other [3-6]. The coded-aperture approach to radiation detection is from the development of the pinhole collimator for X-ray and gamma ray imaging. A multi-pinhole collimator consists of many pinholes, and

improves the signal-to-noise ratio (SNR) maintaining the diameter of the pinholes, which is an important parameter for the resolution of a pinhole imaging system [7].

In the early work on coded apertures, pinholes were randomly distributed on the collimator. The random patterns of pinholes caused difficulties with image reconstruction due to the lack of uniformity in pinhole distribution.

This problem was addressed by the development of uniformly redundant arrays (URAs) and the modified uniformly redundant array (MURA), which is based on pseudo-noise sequences and quadratic residues [8].

In this study, we designed a coded aperture collimator based on a MURA pattern to make a gamma imaging system that can localize a gamma-ray source. We simulated and modeled the coded aperture system that from a radiation source and a collimator to detector. Then

Table 1. Specifications of the simulated detector.

Features	Specifications
Material	CdZnTe
Pixel size	2 mm
Pixel pitch	2 mm
Detector module size	22 × 22 × 5 mm
No. of detector modules	2 × 2 matrix
FOV	44 × 44 mm



Fig. 1. Simulated collimator pattern (MURA).

implement a direct decoding algorithm for coded aperture imaging. In addition, we designed the coded aperture collimator for our gamma imaging detector and did a feasibility test for our coded aperture imaging system.

2. Materials and Methods

2.1 Simulation

We acquired list-mode data sets of a coded aperture simulation with GATE v7.1. The detector consists of 2 × 2 modules which have an 11 × 11 CZT array. Several details are clearly summarized in Table 1.

We modeled MURA coded aperture collimators with various thicknesses. The pixel size of the collimator is determined by the detector pixel. The elements of the coded aperture are 21 × 21 (RANK 11) and 37 × 37 (RANK 19). Locations of opaque and transparent holes in the collimator array are calculated with Eq. (1) [8]:

$$A_{ij} = \begin{cases} 0 & \text{if } i = 0, \\ 1 & \text{if } j = 0, i \neq 0, \\ 1 & \text{if } C_i C_j = +1, \\ 0 & \text{otherwise,} \end{cases} \quad (1)$$

where $C_i = \begin{cases} +1 & \text{if } i \text{ is a quadratic residue modulo } p, \\ -1 & \text{otherwise} \end{cases}$

Fig. 1 shows the two kinds of MURA pattern modeled in our simulation. The RANK 19 collimator has a 2 mm pixel size and uses a 38 mm × 38 mm detector field of view (FOV), while the RANK 11 collimator has a 4 mm pixel size and uses the full FOV of the detector due to the modeled MURA coded aperture collimators with various thicknesses.

2.2 Decoding Algorithm

After getting a photon count distribution from the detector, the source direction can be determined using a




Fig. 2. Block diagram of the direct decoding algorithm in this study.

Table 2. Features of the experimental conditions.

Features	Specifications
Scintillator	CSI, 10 MM THICKNESS
Detector	Hamamatsu, H8500C
Pixel size	2 mm
Pixel pitch	2.065 mm
Pixel number	21 x 21
DAQ system	Vertilon, IQSP418




Fig. 3. Design of the coded aperture collimator (left), the tungsten product (right).

decoding or reconstruction algorithm. We implemented a direct decoding algorithm to estimate location of the source, which is based on second convolution and deconvolution of the coded aperture imaging system.

Our decoding process is shown in the following block diagram. The algorithm makes a decoding matrix for deconvolution and getting the decoding result. The final decoded image is from getting rid of the extended image matrix of the convolution result.

2.3 Experiments

We did experiments to investigate the feasibility of the coded aperture comparison imaging system and to compare the simulation results and the experimental results. We made a set of experiments with our photon detector and scintillator, which can make a high-energy gamma ray into a low-energy photon. The specifications of the detector and scintillator are listed in Table 2.

The 11 × 11 detector pixels that had superior uniformity and resolution to other areas were set as the FOV of the detector to get the experimental data. We designed a RANK 11 coded aperture collimator and made it with tungsten, which is generally used for collimators due to its high density and attenuation coefficient. Fig. 3 shows our coded aperture design and tungsten product.

We used the Co-57 gamma-ray point source, which emits 122 keV photons. The source location was 0° and 9°,



Fig. 4. Experimental imaging module (left), dark box (right).

Location of the source	(0°,0°)	(-9°,0°)	(-5°,0°) + (13°,0°)
Distribution of photon count			
Reconstructed Image			

Fig. 5. Reconstruction images from RANK 11 collimator simulation.




Fig. 6. RMSE and relative sensitivity depending on collimator thickness.

and we simulated the experiments with the same conditions to compare the results. The experiment was performed in a dark box to block out visible light. Fig. 4 shows our experimental imaging modules composed of the detector, the scintillator, the jig, and the collimator.

3. Results

3.1 Simulation Results

Fig. 5 shows decoding images of the simulation results with the RANK 11 coded aperture collimator. Each photon count distribution is listed in the upper line, and reconstructed images are listed in the lower line. The decoded images are 11 x 11 pixels, and pixel size is about 4.4° for a 45° FOV.

	Images at detector plane	FOV images	Pixelated images	Reconstructed images	Simulation results
(0°,0°)					
(-9°,0°)					

Fig. 7. Serial results of experiments.

We ran simulations with RANK 11 and RANK 19 collimators at 2 mm to 10 mm thicknesses (at 2 mm intervals) to evaluate the performance of the collimators. Fig. 6 shows the root mean square error (RMSE) of the reconstructed image and relative sensitivity to the 2 mm thick collimator.

3.2 Experimental Results

Fig. 7 shows the experimental results from the photon count distribution at the detector plane, FOV image, and 11 x 11 pixelated image, compared to the reconstructed image. The full width half maximum (FWHM) of the 0° source was 5.09°, and the 9° source was 4.82° in the reconstructed image.

4. Discussion and Conclusion

In this study, we implemented a previous coded aperture imaging system basically from simulation to experiment. In Fig. 5, our decoding process creates a source localization image that matches the actual location of sources. In experimental results, Fig. 7 shows reliable results compared to the simulation results. We will continue to improve the decoding algorithm, and will optimize the collimator for high-energy gamma rays emitted from nuclear power plants.

Acknowledgement

This work was supported by the Korea Institute of Energy Technology Evaluation and Planning (KETEP) and the Ministry of Trade, Industry & Energy (MOTIE) of the Republic of Korea (No. 20161520302180) and by a National Research Foundation of Korea (NRF) grant funded by the Korean Government (MSIP) (No. 2016R1A2B2007551).

References


[1] Y. Shao et al., "Development of a PET Detector System Compatible with MRI/NMR Systems", *IEEE Trans. Nucl. Sci.*, Vol. 44, No. 3, 1167-1171, 1997. [Article \(CrossRef Link\)](#)

[2] Joshi, Sonal. "Coded Aperture Imaging Applied to Pixelated CdZnTe Detectors", *PhD diss.*, University

- of Michigan, 2014. [Article \(CrossRef Link\)](#)
- [3] Tumerl, T. et al., "All-sky x-ray & gamma-ray astronomy monitor (AXGAM)", *IEEE Trans. Nucl. Sci.*, 44 (3), 572-576, 1997. [Article \(CrossRef Link\)](#)
- [4] Gehrels et al., "The swift gamma ray burst mission." *Astrophysical J.*, 611 (2), 1005-1020, 2004. [Article \(CrossRef Link\)](#)
- [5] Hong et al., "Laboratory coded aperture imaging experiments: radial hole coded masks and depth-sensitive CZT detectors", In: *Proceedings SPIE, Hard XRay and Gamma-Ray Detector Physics VI*, 5540, pp. 1-10, 2004. [Article \(CrossRef Link\)](#)
- [6] Del Monte et al., "An X-ray imager based on silicon microstrip detector and coded mask", *Nucl. Instrum. Methods Phys. Res. Sect. A*, 576 (1), 191-193, 2007. [Article \(CrossRef Link\)](#)
- [7] Dicke, "Scatter-hole cameras for x-rays and gamma rays", *Astrophysical J.*, 153, 101-106, 1968. [Article \(CrossRef Link\)](#)
- [8] Gottesman et al., "New family of binary arrays for coded aperture imaging", *Appl. Opt.*, 28 (20), 4344-4352, 1989. [Article \(CrossRef Link\)](#)



Kwangdon Kim received his BS degree in Radiological Science from Korea University in 2013. From 2011 he was a student research member of Medical Information Processing Laboratory (MIPL) in Korea University. He is now on MS and PhD course in Department of IT convergence, Korea University, Seoul, Korea and working in MIPL



Hakjae Lee received the B.S degree electric and electronic engineering from the Chung-Ang University in 2005 and the M.S. and Ph.D. degree in Bio-convergence Engineering from the Korea University at Seoul, Korea in 2010 and 2014. Now he is working for research institute of Global Health Tech. at Korea University as a research professor. His research interests include medical image processing and radiation detector development for medical applications.



Kisung Lee received BS and MS degrees in electronics engineering from the Korea University in 1990 and 1992, respectively, and a PhD degree in electrical engineering from the University of Washington at Seattle in 2003. During 2005 to 2007, he was an assistant professor at the Kongju National University, Korea. He has served as a reviewer of IEEE Transaction on Nuclear Science and IET image Processing. He served as a Treasurer of 2013 IEEE Science Symposium, Medical Imaging Conference and Workshop on Room-Temperature. He is now with the Department of Radiologic Science at the Korea University and has worked in the area of radiation detectors, physics in x-ray and gamma-ray imaging systems, and image processing algorithms for medical applications.



Jinwook Jang received his BS degree in Bio-medical Engineering from Korea University in 2015. From 2014 he was a student research member of Medical Information Processing Laboratory (MIPL) in Korea University. He is now on MS course in Bio-convergence Engineering, Korea University and working in MIPL



Yong Hyun Chung is a Professor of department of Radiation Convergence Engineering at Yonsei University from 2006. He received his B.S., M.S., and Ph. D degree in department of nuclear and quantum engineering in KAIST, South Korea, and majored in the radiation detection and medical imaging especially for nuclear medicine instrumentations. From 2002 to 2006, he worked as a Researcher at Department of Nuclear Medicine, Samsung Medical Center, Seoul, Korea, at Center for Clinical Research, Samsung Biomedical Research Institute, Seoul, Korea and at Crump Institute for Molecular Imaging, David Geffen School of Medicine, UCLA, Los Angeles, USA. His current interests lie in the area of non-invasive imaging techniques, like gamma camera, single photon emission computed tomography (SPECT), positron emission tomography (PET) and specific nuclear material (SNM) monitoring system.



Jinhun Joung went on to receive his MSE and PhD degrees in Biomedical Engineering and Electrical Engineering from the University of Southern California in 1997 and the University of Washington in 2001, respectively. He then began his industrial career with the Molecular Imaging Group,

Siemens Medical Solutions, USA. While there he held positions as Principle Research Scientist, Sr. Principle Research Scientist, Patent Manager and Sr. Project Manager from 2001 to 2009. He served as elected member of NMISC from 2011 to 2013, IEEE NSS/MIC scholarship chair in 2014. Dr. Joung is currently the Chief Executive Officer of NuCare Medical Systems, Inc. and an Adjunct Professor of Biomedical Engineer at Korea University. He has involved himself in various research and development projects in academic and industrial settings related to “advanced nuclear medicine detector technology”. His research focus is “multi-modality molecular imaging detector technology” for both in-vivo diagnostic and therapeutic imaging procedures.